

LV5380 Software Upgrade Procedure

Caution: Please DO NOT turn off the power during software upgrade.

- 1. Unzip the file LV5380.zip
- 2. Copy the "LV5380" folder to the USB thumb drive from the PC. (The LV5380 folder includes the SYS and LDR subfolders)
- 3. Plug in the USB thumb drive to the LV5380.
- 4. Power on the LV5380.
- 5. Press the **SYS** button on the front panel.
- 6. Press [F5] NEXT MENU.
- 7. Press [F1] INIT
- 8. Press [F5] FIRMWARE UPDATE.
- 9. Press [F3] YES
- 10. Update will take few minutes and it will reboot after update is completed.

Phone: +1 (714) 487-5200

11. END

LEADER INSTRUMENTS CORPORATIONToll Free Phone: +1 (800) 645-5104

Email: Service@LeaderAmerica.com

Service Department Fax: +1 (714) 487-5201

Web: www.LeaderAmerica.com